

Studies shows critically ill patients use Masimo ORi to reduce hyperoxia

27 August 2019 | News

Researchers evaluated with pure oxygen saturation (SpO₂ comparison) monitoring, Masimo ORi™ (oxygen reserve index) monitoring critically ill patients in the possibility of shortening the time of high oxygen.



Masimo, on 27 August 2019, announced the latest publication of a report on 150 randomized controlled trials of ICU patients in Intensive Care Medicine. Researchers evaluated with pure oxygen saturation (SpO₂ comparison) monitoring, Masimo ORi™ (oxygen reserve index) monitoring critically ill patients in the possibility of shortening the time of high oxygen. ORi is an index of the degree of oxygenation in the moderate hyperoxia section, defined as the oxygen artery partial pressure PaO₂ ranging from 100 to 200 mm Hg. ORi is a graded "index" between 0.0 and 1.0, and clinicians can use their trends to understand changes in the patient's oxygen reserve.

Dr Sigismond Lasocki, Ph.D., CHU Angers, France, and colleagues sought to assess whether the inhaled oxygen fraction (FiO₂) in patients with mechanical ventilation can be moderately high by monitoring ORi rather than simply SpO₂. Time of oxygen state (PaO₂ ≥ 100 mm Hg). 150 adult patients with critical illness had received mechanical ventilation for at least 2 days and were randomized to ORi and control groups. Root using Masimo® care patients and interworking with internet Rainbow® sensors and measuring ORi SpO₂. In the ORi group, the nurses lowered FiO₂ when ORi ≤ 0.01. In the control group (only SpO₂ was monitored), nurses lowered FiO₂ when SpO₂ ≥ a certain upper limit of prescription. PaO₂ was measured by analyzing arterial blood gas.

The analysis data were from 75 patients in the ORi group and 71 patients in the control group. In the statistical sense, the researchers found that monitoring ORi significantly reduced the percentage of days in which patients were exposed to high oxygen without significantly increasing hypoxia: hyperoxia, 14% (0 to 33%) vs. control group in the ORi group. 33% (11 to 56%) (p=0.003). The percentage of hours of hyperoxia in the ORi group was also "lower": PaO₂ ≥ 100 mm Hg, 7.4% (0–24.8%) in the ORi group and 17.3% in the control group (3.8–43.1%) (p=0.0069), PaO₂ ≥ 120 mm Hg, 0% (0–7.2%) in the ORi group and 5.6% (0–18.1%) in the control group (p=0.0037).

The researchers concluded that " using ORI monitoring to titrate oxygen supply rates can shorten the patient's high oxygen rate compared to SpO₂ alone, probably because nurses are reluctant to lower the oxygen supply rate when SpO₂ is in the normal range. The department has implemented a nurse-driven procedure that regulates FiO₂ by SpO₂, which explains why the percentage of hyperoxia time observed in the control group is much lower than the usual report (30 vs 60%). In the department that regulates the oxygen supply rate protocol, this strategy of lowering the oxygen supply rate by ORi (higher PaO₂ can be detected) may be more efficient. SpO₂ can still be used as a warning of hypoxia, and ORi can be used as a high Oxygen warning. A larger sample size study is needed to assess the clinical benefit of the strategy."

Ori has not been approved by the FDA 510(k) and is not available for sale in the United States.